

Datum izdavanja dodatka: 05.05.2022.

Issue date of annex:

Zamjenjuje dodatak od:

Replaces Annex dated:

Dodatak Sertifikatu o akreditaciji sa akreditacionim brojem broj: L_K 14.02
Annex to Accreditation Certificate - Accreditation Number: L_K 14.02

Standard: MEST EN ISO/IEC 17025:2018

Datum dodjele/ obnavljanja akreditacije:

04.03.2014. / 05.05.2022.

Date of granting/ renewal of accreditation.

Akreditacija važi do: 04.05.2026.

Accreditation is valid until:

Akreditovana laboratorija za kalibraciju

Accredited calibration laboratory

Zavod za metrologiju
Sektor za metrološku sljedivost i državne etalone
Arsenija Boljevića b.b. Podgorica

Bureau of metrology
Sector of metrological traceability and national measurement standards
Arsenija Boljevića b.b. Podgorica

Područje akreditacije / Scope of accreditation

Kalibracija tegova, tegova slobodnih nazivnih masa i neautomatskih vaga

Kalibracija mjerila temperature i relativne vlažnosti

Kalibracija mjerila zapremine od stakla i mjerila zapremine sa klipom

Kalibracija etalona prelivnih pipeta i mjernih posuda

Kalibracija etalona i mjernih uređaja dužine

Kalibracija mjerila pritiska

Kalibracija etalona i mjerila električnih veličina

Kalibracija mjerila frekvencije i vremenskog intervala

Kalibracija mjerila koncentracije radona u vazduhu

Calibration of weights, weights with free nominal masses and non-automatic weighing instruments

Calibration of temperature and relative humidity gauges

Calibration of laboratory glassware and piston-operated volumetric apparatus

Calibration of etalons over flow pipettes and standard capacity measures and calibration of standard capacity measures

Calibration of standards and instruments for measurement of length

Calibration of pressure gauges

Calibration of standards and instruments for measurement of electrical quantities

Calibration of instruments for measurement of frequency and time interval

Calibration of measuring devices for measurement concentration of radon in the air

Datum izdavanja dodatka: 05.05.2022.

Issue date of annex:

Zamjenjuje dodatak od:

Replaces Annex dated:

Detaljan obim akreditacije / Detailed scope of accreditation

Laboratorija za jonizujuća zračenja, Arsenija Boljevića b.b. Podgorica

* Metoda kalibracije se sprovodi na terenu

** Metoda kalibracije se sprovodi u Laboratoriji i na terenu

Red br. No	Oblast kalibracije <i>Calibration field</i>	Mjerna veličina/predmet kalibracije <i>Measurand / Calibration object</i>	Mjerno područje <i>Measurement range</i>	Mogućnost mjerenja i kalibracije ¹ <i>Calibration and Measurement Capability</i>	Referentni dokument <i>Reference document</i>	Oznaka Lokacije <i>Location</i>
1.	Jonizujuća zračenja <i>Ionising radiation</i>	Koncentracija radona u vazduhu <i>Radon concentration in the air</i>	do/up to 1000 Bq/m ³	11,9%	Metoda direktnog poređenja <i>Direct comparison method</i> QP.7.2/03-LIR Izdanje/Issue 01/02, 22.04.2020.	L1
			1000 Bq/m ³ do/to 3000 Bq/m ³	8,7%		
			3000 Bq/m ³ do/to 10000 Bq/m ³	5,2%		
			10000 Bq/m ³ do/to 20000 Bq/m ³	5,6%		

Oznaka lokacije <i>Location code</i>	Detalji o lokaciji (naziv i adresa) <i>Location details (title and adress)</i>
L1	Arsenija Boljevića b.b. Podgorica

Legenda / Legend

Skraćena oznaka referentnog dokumenta <i>Abbreviation of reference document</i>	Naziv metode/referenca <i>Title of method/reference</i>
QP.7.2/03-LIR	Tehnička procedura za kalibraciju uređaja za mjerenje koncentracije radona u vazduhu / IEC 61577-1,2,3,4,5